◆ ZERTIFIKAT ◆ CERTIFICATE ◆ CERTIFICAT ◆ CERTIFICADO ◆ CERTIFICATO ◆ CERTIFICAAT ◆ СВИДЕТЕЛЬСТВО ◆ 認証証書 ◆

Certificate

No: AL 96 02 17955 006

TÜV PRODUCT SERVICE

Omicron Electronics GmbH

Brolsstr. 39
A - 6844 Altach

with production facilities
17955

is authorized to label the following products with the
certification mark A,B or L
as shown in the certification mark list. See also notes overleaf.

Product: **Netzgeräte**
(Current Amplifier)

Model: **Typ / type CMA 56 und/and CMA 156**
Weitere Angaben siehe Anlage 1
For further information see attachment 1

Parameters:

	CMA 56 A	CMA 56 B	CMA 156 C
Nennspannung: (Rated input voltage)	110-240 V	220-240 V	110-240 V
Nennleistung: (Rated input power)	1000 VA	1000 VA	1000 VA
Nennfrequenz: (Rated frequency)	50/60 Hz	50 Hz	50/60 Hz
Schutzklasse: Protection class)	I	I	I

The product meets the requirements of the Equipment Safety Law and was tested according to:

EN 61010-1: 1993

Report No: 01210601501, E - (AL 95 04 17955 005)

Released with the above certificate number by the certification body of TÜV PRODUCT SERVICE GMBH.

Organization unit: 1ITE / abs
Date: 03-27-1996
Translation of the German original.