

◆ ZERTIFIKAT ◆ ZERTIFIKAT ◆ ZERTIFIKAT ◆ ZERTIFIKAT ◆ ZERTIFIKAT ◆ ZERTIFIKAT ◆ ZERTIFIKAT ◆ ZERTIFIKAT ◆ ZERTIFIKAT ◆ ZERTIFIKAT ◆
 ◆ CERTIFICATE ◆ CERTIFICATE ◆ CERTIFICATE ◆ CERTIFICATE ◆ CERTIFICATE ◆ CERTIFICATE ◆ CERTIFICATE ◆ CERTIFICATE ◆ CERTIFICATE ◆ CERTIFICATE ◆
 ◆ CERTIFICADO ◆ CERTIFICADO ◆ CERTIFICADO ◆ CERTIFICADO ◆ CERTIFICADO ◆ CERTIFICADO ◆ CERTIFICADO ◆ CERTIFICADO ◆ CERTIFICADO ◆ CERTIFICADO ◆
 ◆ CERTIFICATO ◆ CERTIFICATO ◆ CERTIFICATO ◆ CERTIFICATO ◆ CERTIFICATO ◆ CERTIFICATO ◆ CERTIFICATO ◆ CERTIFICATO ◆ CERTIFICATO ◆ CERTIFICATO ◆
 ◆ CERTIFICAAT ◆ CERTIFICAAT ◆ CERTIFICAAT ◆ CERTIFICAAT ◆ CERTIFICAAT ◆ CERTIFICAAT ◆ CERTIFICAAT ◆ CERTIFICAAT ◆ CERTIFICAAT ◆ CERTIFICAAT ◆
 ◆ СВИДЕТЕЛЬСТВО ◆ СВИДЕТЕЛЬСТВО ◆ СВИДЕТЕЛЬСТВО ◆ СВИДЕТЕЛЬСТВО ◆ СВИДЕТЕЛЬСТВО ◆ СВИДЕТЕЛЬСТВО ◆ СВИДЕТЕЛЬСТВО ◆ СВИДЕТЕЛЬСТВО ◆ СВИДЕТЕЛЬСТВО ◆ СВИДЕТЕЛЬСТВО ◆
 ◆ 認証証書 ◆ 認証証書 ◆ 認証証書 ◆ 認証証書 ◆ 認証証書 ◆ 認証証書 ◆ 認証証書 ◆ 認証証書 ◆ 認証証書 ◆ 認証証書 ◆

Certificate



No: AL 96 03 17955 007

Omicron Electronics GmbH

Brolsstr. 39
A - 6844 Altach

with production facilities
17955

is authorized to label the following products with the
certification mark A,B or L
as shown in the certification mark list. See also notes overleaf.

Product: Netzgeräte
(Prüfgerät/Testequipment)

Model: Typ / type CMC 151 und/and CMC 156
Weitere Angaben siehe Anlage 1
For further information see attachment 1

Parameters:	CMC 151	CMC 156
Nennspannung: (Rated input voltage)	110-240 V	110-240 V
Nennleistung: (Rated input power)	600 VA	600 VA
Nennfrequenz: (Rated frequency)	50/60 Hz	50/60 Hz
Schutzklasse: Protection class)	I	I

The product meets the requirements of the Equipment Safety Law and
was tested according to:

EN 61010-1: 1993

Report No: 01230500801

Released with the above certificate number by the
certification body of TÜV PRODUCT SERVICE GMBH.

Organization unit: 1ITE / abs
Date: 03-27-1996

Translation of the German original.

